



Table of contents

1.	Introduction	
2.	Technology description 2.1 About PEA™ 2.2 Quality Control	
3.	Product description 3.1 Reagents supplied in Olink kits	
4.	Additional requirements 4.1 NPX [™] Signature 4.2 Required consumables (not supplied) 4.3 Required equipment (not supplied)	12
5.	Laboratory work safety and guidelines	14 14 14 16
6.	 Preparation 6.1 Experimental design. 6.2 Create Olink extension and pre-amplification program 6.3 Bench setup 6.4 Time indications and limits. 6.5 Sample preparation. 6.6 Plate layout 	18
7.	Operation7.1Overview7.2Incubation reaction step7.3Prepare bench7.4Incubation instruction7.5Extension and amplification step7.6Detection step	22 22 22 22 22 22 22 22 22 22
8.	 StandardBioTools® BioMark[™] system	3(3(3(32
9.	Other resources 9.1 Olink FAQs 9.2 Data analysis troubleshooting 9.3 Lab instruction video 9.4 PEA technology video	
10.	Revision history	

1. Introduction

1.1 Intended use

Olink high-multiplex immunoassay panels provide an efficient and innovative tool for targeted human protein biomarker discovery, development and validation.

The Olink[®] Target 48 panels enable rapid analysis with minimal sample consumption and exceptional data quality, with protein concentrations reported in standard units (pg/mL) as default and optionally in relative (NPXTM) units. Using just 1 μ L of sample, 45 proteins are assayed simultaneously on each panel, and results are generated within 24 hours.

1.2 About this manual

The Olink[®] Target 48 User Manual provides you with the instructions needed to run an Olink[®] Target 48 panel using FORMULATRIX[®] F.A.S.T.[™] and Olink[®] Signature Q100. If using the BioMark[™] system, refer to *8. StandardBioTools[®] BioMark[™] system.*

1.2.1 Definition of alert levels

🕂 WARNING: Indicates a potentially hazardous situation which, if not avoided, could result in injury.

IMPORTANT: Indicates an important action that may impair the results if not performed correctly.

NOTE: Contains information that can make it easier to understand or perform a certain task.

SAFE STOPPING POINT: Indicates a step where the protocol can be safely paused and restarted at a later time.

TIME SENSITIVE STEP: Indicates a step that must be performed within a limited time period. Results may be impaired if not performed correctly.

1.3 Important user information

1.3.1 Sample preparation

To reduce sample-handling time, samples should be distributed in 8-well strips or on a 96-well plate prior to the start of the run.

1.3.2 Sample types

Olink[®] Target 48 panels have been validated using EDTA plasma and serum samples. Different sample matrices are expected to affect the detection of specific proteins in different ways. Extreme levels of IgG or fluorescent particles can interfere with the Olink assay. For more information on sample types, please find the Data Validation documents corresponding to each panel on our website, or contact Olink support at support@olink.com.

1.3.3 Pipettes

Maintain and calibrate the pipettes regularly. The F.A.S.T. should be covered by an active service contract with regular preventative maintenance.

1.4 Associated documentation

1.4.1 Olink documentation

The Olink documents are available from the Olink website: www.olink.com/downloads.

User Manuals

- Olink[®] Signature Q100 User Manual
- NPX[™] Signature User Manual

White papers

- "Strategies for design of protein biomarker studies"
- "Data normalization and standardization"

Material Safety Data Sheet

- Panel-specific Detection kit
- Panel-specific Cytokine Control kit
- Panel-specific Cytokine Probe kit

1.4.2 Other documents

For information on the Standard Biotools IFC Controller HX and BioMark System, read the following User Guides that can be found on <u>www.standardbio.com</u>:

- Standard Biotools® IFC Controller User Guide PN 68000112
- Standard Biotools® Real-Time PCR Analysis User Guide PN 68000088
- Standard Biotools® Data Collection Software User Guide PN 68000127

For information on the MixMate® read the Operating Manual that can be found on <u>www.eppendorf.com</u>:

• MixMate® Operating Manual - 5353 900.058-00/092018

For information on the FORMULATRIX F.A.S.T., read the Operating Manual that can be found on <u>https://help.formulatrix.com/fast/</u>:

• F.A.S.T. Web Help

1.5 Technical support

For technical support, please contact Olink Proteomics at support@olink.com.

2. Technology description

2.1 About PEA[™]

The technology behind the Olink panels is called Proximity Extension[™] Assay (PEA[™]). The three core steps of the technology are illustrated below.



1. Incubation step

Duration: Overnight 16-22 hours

The 45 antibody pairs, labelled with DNA oligonucleotides, bind to their respective protein in the samples.

2. Extension and amplification

Duration: 2 hours

Oligonucleotides that are brought into proximity hybridize, and are extended using a DNA polymerase. This newly created piece of DNA barcode is amplified by PCR.

3. Detection

Duration: 3 hours

The amount of each DNA barcode is quantified by microfluidic qPCR.

2.2 Quality Control

Olink has developed a built-in Quality Control (QC) system using internal controls, for its multiplex biomarker panels. This system enables full control over the technical performance of assays and samples.

2.2.1 Internal controls

The QC system consists of three internal controls that are spiked into every sample and designed to monitor the three main steps of the Olink protocol: Immuno reaction (incubation step), extension and amplification/detection.



Incubation Control: The Incubation Control is a non-human antigen measured with PEA. This control monitors potential technical variation in all three steps of the reaction.

Extension Control: The Extension Control is composed of an antibody coupled to a unique pair of DNA-tags. These DNA-tags are always in proximity, so that this control is expected to give a constant signal independently of the immuno reaction. This control monitors variation in the extension and amplification/detection step and is used in the normalization process of each sample.

Detection Control: The Detection Control is a complete double stranded DNA amplicon which does not require any proximity binding or extension step to generate a signal. This control monitors the amplification/detection step.

2.2.2 External controls

There are three required external controls that are added in triplicate or duplicates on each sample plate. On the illustrated plate there are 40 samples, three Sample Controls, two Negative Controls and three Calibrators.



Sample Control: The Sample Control is a pooled human plasma sample, with spiked in recombinant antigens for some proteins to ensure that all 45 proteins are detected, and included in triplicate on each plate. The Sample Control is used to assess potential variation between runs and plates, as well as quality control of the quantification of data in pg/mL. See the Quantified data section below. Validation data for all panels is available on the Olink website. An example of inter- and intra-assay CV is illustrated here:



Negative Control: Negative Control is included in duplicate on each plate and consists of buffer. These are used to monitor any background noise generated when DNA-tags come in close proximity without prior binding to the appropriate protein. The Negative Controls set the background levels for each protein assay and are used to calculate the limit of detection (LOD) for each plate, defined as three fixed standard deviations above average for the Negative Controls.

Calibrator: The Calibrator is a pooled human plasma sample with spiked in recombinant antigens for proteins with low endogenous levels to ensure that all 45 proteins are detected within LOQ. The calibrator is used to adjust the predefined standard curve along the y-axis during development at Olink. The calibrator is also included in triplicate on each sample plate and used to normalize all samples. See the Quantified data section below for details.

2.2.3 Data analysis

Protein expression is reported in standard concentration units (pg/mL) as default. Optionally, it can also be reported in relative Normalized Protein eXpression (NPX) units. Quality control and normalization is achieved using the Olink NPX Signature software, refer to *4.1 NPXTM Signature*.

2.2.4 Results in relative unit (NPX)

Olink reports protein expression levels using an arbitrary unit called Normalized Protein eXpression (NPX). NPX is a relative quantification unit on log2 scale that allows users to identify changes for individual protein levels across their sample set, and then use this data to establish protein signatures. NPX is derived from the Ct values obtained from the qPCR using the following equations for Olink Target 48 panels:

Each sample is normalized with its corresponding extension control level:

Each plate is normalized with its Calibrators to generate NPX levels and adjusted against a pre-determined bridging factor.

Bridging factor Assay j - (dCt_{Sample i, Assay j} - Median(Calibrator_{Sample i, Assay j})) = NPX_{Sample i, Assay j}

NOTE: The NPX value will differ between the same protein measured on Olink Target 96 and Olink Target 48. This is mainly due to the fact that Target 48 uses a bridging factor and Target 96 a correction factor, and the sample used for plate normalization differs.

The NPX unit is unique to each protein assay, meaning that even if two different proteins have the same NPX values, their concentrations in pg/mL may differ.

2.2.5 Results in standard unit (pg/mL)

The method for quantification of data in standard units (pg/mL) for Olink Target 48 is described in this section.

Method overview

The method requires fewer sample wells on each plate/run for controls compared to many other platforms, making sure that as many wells as possible can be used for customer samples.

Before run (during product development at Olink)

- 1. A precise pre-defined standard curve is established for each protein.
- 2. The triplicate calibrator is used to adjust the standard curve along the y-axis to normalize between plates and a second normalization, utilizing a bridging factor, is applied to ensure accuracy between production batches.
- 3. A 4-PL model fit is performed to define the standard curve mathematically in the measurement range for each protein in the panel.

At each run

- 1. The median value of the calibrator triplicates is used to adjust the measured samples along the y-axis to normalize between plates. The samples are also adjusted against the bridging factor.
- 2. The measured sample values are related back to the adjusted standard curve model which is used to translate the measured NPX value to the protein concentration in pg/mL.

Standard curve

During development, a thorough fine-tuned 32-point standard curve is developed for each protein biomarker simultaneously including all 45 proteins. Multichannel pipetting and numerous replicates of the curves are used to minimize errors and establish an accurate immunoassay curve fitting. Due to the wide dynamic range of the PEA assays in the panel, a large spread of measurement points are required to cover the entire range. Eight measurement points, which is commonly used in many technologies, would not suffice for Olink's multiplex measurement of 45 assays. The predefined

standard curves avoid operator dependent reconstitution and pipetting of standard curves at each lab and for each run. The figure below shows an example of a point standard curve defined for each assay during development. Standard curves for each assay can be found via the panel product page (www.olink.com/t48cyt).



4PL model

A 4PL model curve fit is used to describe the immunoassay standard curve, indicated by the turquoise line in the figure above. When running a project, the measured patient sample value (represented by the dotted blue arrows in the figure) is related back to the adjusted standard curve model which translates the measured value to the protein concentration in pg/mL. Repeated testing and validation show that the 4PL curve fitting describes the standard curve well, and can be used to correctly estimate the concentration in analyzed samples within the limits of quantification (LOQ). The lower and upper limits of quantification (LLOQ and ULOQ) are defined during the development of the panel, see the panel specific Validation data document available on the Olink website (www.olink.com) for details.

QC of assays in pg/mL

The QC of each assay is performed utilizing the Sample Controls. They are assessed for both accuracy and precision (CV%) for each individual assay to confirm that the concentration values in pg/mL are within the expected limits (\pm 30% of expected value) and assure the correct functionality of the quantification method.

3. Product description

3.1 Reagents supplied in Olink kits

Each Olink kit contains reagents for 48 wells, sufficient for 40 samples and 8 controls. The reagents are supplied in three individual boxes. Storage temperature and expiry date for the components are stated on the label on each box.

NOTE: All reagents for Olink Target 48 kits are lot specific and reagents from different kit lots cannot be combined. It is also not possible to use reagents from Olink Target 96 kits with reagents from the Olink Target 48 kits.

3.1.1 Probe kit

The Olink Probe kit should be stored at +4 $^{\circ}$ C. The Probe kit includes:

Part	Description
Olink® 1-48 plex Incubation Solution	Contains components needed for the immuno reaction
Olink [®] Target 48 Frw-probes	Contains 46 antibody probes labeled with Frw-oligos
Olink [®] Target 48 Rev-probes	Contains 46 antibody probes labeled with Rev-oligos
Olink [®] Target 18 Cyteking	



Product no: 93200A Lot no: B01010A Expiry date: DEC 2023



3.1.2 Detection kit

The large box contains the Olink Detection kit. It should be stored at -20 °C and includes:

Part	Description
Olink® 1-48 plex PEA Solution	Contains components needed for the extension reaction
Olink [®] 1-48 plex PEA Enzyme	For extension of Frw- and Rev-probes bound to their target
Olink [®] 1-48 plex PCR Polymerase	For pre-amplification of the extension product created by the PEA Enzyme
Olink® 1-48 plex Detection Solution	Contains components needed for the detection reaction
Olink [®] 1-48 plex Detection Enzyme	For qPCR amplification
Olink® 1-48 plex Primer Plate	96-well plate with ready-to-use primers for amplification of extension product
Olink [®] 1-48 plex Negative Control	For determination of background levels
Olink [®] 1-48 plex PEA Enhancer	Contains components needed for the extension reaction

Olink[®] Target 48 Cytokine Detection kit Product no: 93200B Lot no: B01010B Expiry date: DEC 2023



3.1.3 Control kit

The Olink Control kit should be stored at -80 °C. The Control kit includes:

Part	Description
Olink [®] Target 48 Calibrator	Used for normalization between plates and of samples
Olink [®] Target 48 Sample Control	Used for quality control of quantification of data in pg/mL and to assess variation between runs and plates

Olink[®] Target 48 Cytokine Control kit Product no: 93200C Lot no: B01010C Expiry date: DEC 2023



4. Additional requirements

4.1 NPX[™] Signature

NPX[™] Signature software is an easy to use data import and pre-processing tool developed by Olink Proteomics. The software lets you import data, validate data quality and normalize Olink data for subsequent statistical analysis. See the Olink[®] NPX Signature User Guide for more information.

4.2 Required consumables (not supplied)

- Pipette filter tips
- Microcentrifuge tubes (1-1.5 mL)
- F.A.S.T. Disposable Pipette Tips (FORMULATRIX® catalogue number 233589)
- Centrifuge tube (> 11 mL)
- 8-well strips with lids
- 96-well full skirted PCR plate (à 0.2 mL)
- 96-well semi skirted PCR plate (à 0.2 mL)
- Multichannel pipette reservoir
- Adhesive plastic film (heat-resistant)
- High purity water (sterile filtered, MilliQ[®] or similar)
- Olink[®] 48.48 IFC for Protein Expression

Contact Olink support at support@olink.com for specific recommendations.

4.3 Required equipment (not supplied)

- FORMULATRIX F.A.S.T. running software v1.27.0 or later
- FORMULATRIX F.A.S.T. 96 well plate adapter
- FORMULATRIX F.A.S.T. strip tube adapter
- Vortex
- MixMate
- 96 tube holder for MixMate
- Centrifuge for plates
- Microcentrifuge for tubes
- Freezing block (-20 °C) for enzyme handling
- Thermal cycler with:
 - Heated lid
 - Temperature range from +50 °C to +95 °C
 - Validated for 0.1 mL volumes (important)
 - 96-well format (recommended)
- Refrigerator or cold room (+2 °C to +8 °C) where the light is off when the door is closed
- Freezer (-20 °C) where the light is off when the door is closed
- Freezer (-80 $^\circ\text{C})$ where the light is off when the door is closed

- Olink Signature Q100
 - alternatively StandardBioTools[™] Biomark or Biomark[™] HD system

5. Laboratory work safety and guidelines

5.1 Safety considerations

5.1.1 Safety

Follow general laboratory safety procedures:

- Use gloves, safety goggles and protective clothing when performing the experiments.
- Handle solutions with particular caution, as Dimethyl sulfoxide (DMSO) is known to facilitate the entry of organic molecules into tissues.
- Handle and dispose of hazardous sample material according to local regulations.

For complete safety information regarding Olink kits and material, refer to Safety Data Sheets (SDS) available on the Olink website: <u>www.olink.com/downloads</u>.

5.1.2 PCR technology

PCR technology is sensitive to contaminations. Perform the Detection step in a PCR room, separate from the room where the Immuno and Extension steps are performed. Maintain and calibrate the PCR and Olink Signature Q100 instruments regularly.



5.2 Pipetting guidelines

5.2.1 Tips and tricks for pipetting

- Calibrate pipettes regularly (at least a 6 month interval).
- Pipette near the liquid surface.
- Let the reagents and liquids reach room temperature before use to maximize accuracy.
- Do not turn the pipette on its side when there is liquid in the tip as liquid might contaminate the interior of the pipette.
- Keep pipettes vertical while pipetting and pipette to the bottom of the wells, unless otherwise stated in the protocol.

5.2.2 Pipetting techniques

Both forward and reverse pipetting is used in the Olink protocol.

Forward pipetting

Forward pipetting is the most commonly used pipetting technique and is performed as described in the following instruction:

NOTE: Ensure that all tips contain the exact same volume if a multichannel pipette is used.

- 1. Press the operating button to the first stop.
- 2. Dip the tip into the liquid just below the liquid surface, and slowly release the operating button all the way up.
- 3. Press the operating button to the first stop to dispense the liquid into the receiving vessel and then press the operating button to the second stop. This action will empty the tip. Remove the tip from the vessel.
- 4. Release the operating button all the way up and discard the empty pipette tip.



Reverse pipetting

Reverse pipetting improves precision with smaller volumes and viscous solutions, and is performed as follows:

NOTE: Ensure that all tips contain the exact same volume if a multichannel pipette is used.

- 1. Before filling the pipette with liquid, push the operating button past the first stop and then dip the tip into the liquid, just below the liquid surface.
- 2. Slowly release the operating button all the way up. This will fill the pipette tip with a volume larger than the set volume.
- 3. Press the operating button to the first stop to dispense the liquid into the recieving vessel. This volume is equal to the set volume, and a small volume will remain in the pipette tip. Hold the button pressed down to the first stop if the same pipette tip is going to be used to transfer the same liquid to another well.
- 4. Dip the pipette tip in the liquid again, and slowly release the operator button all the way up.
- Continue pipetting into the recieving vessel.
 Repeat steps 4 and 5 as shown in the illustration, until liquid has been transferred into all wells.
- X. Discard the small volume left in the pipette tip once the pipetting of the liquid is done. Release the operating button all the way up and discard the empty pipette tip.

	1	2	3	4	5	 Х	En	ıd
First stop		\uparrow	\checkmark	\uparrow	\checkmark		\uparrow	•
Second stop	\checkmark					\downarrow		

5.3 Vortexing guidelines

Correct vortexing is critical when running Olink panels.

5.3.1 Vortexing instruction

Correct vortexing is essential for generating reproducible results. To vortex thoroughly, follow these steps:

Vortexing using MixMate

- The MixMate should be used for the vortexing of PCR-plates in the pre-PCR steps. For the Detection step it is recommended but not required.
- Start with inserting the 96 tube holder in the plate holder:
 - Hold the 96 tube holder up against the back edge of the plate holder.
 - Engage the 96 tube holder by pressing gently on the front.
- Insert the plate into the tube holder and press the PCR plate into the bores of the tube holder and make sure that the plate is seated evenly. Note that misplacement of the plate will result in uneven mixing of the wells and low-quality data.
- Set mixing speed at 2000 rpm and mixing time at 30 seconds.
- Start mixing by pressing the start/stop key. After 30 seconds, MixMate will automatically stop.
- Take out the plate by pulling it out from the front side of the plate holder.
- For fully skirted plates the MixMate should be operated without the plate holder. Please see the MixMate Operating Manual for further instructions.

Manual vortexing

A manual vortex is needed to vortex all separate tubes throughout the protocol. A manual vortex may also be used for the vortexing of plates during the detection step, if so, follow the below instructions. You can also see our vortexing <u>video</u> for further guidance on maual vortexing.

- When using a manual vortex, cover the rubber platform of the vortex with adhesive plastic film to make it easier to slide the 96-well plate during vortexing.
- Vortex for 20-30 seconds at full speed.
- Visually inspect the wells during vortexing to ensure complete mixing. The liquid should swirl in the wells.
 - 1. Move the plate back and forth over the vortex in horizontal direction. Make sure that the wells at the outside edges of the plate (columns 1 and 12) are also vortexed.



2. Tilt the plate away from you and move it back and forth over the vortex.



3. Tilt the plate towards you and move it back and forth over the vortex.



4. Turn the plate 180° and repeat steps 1-3.

5.3.2 Sterile lab environment

The high sensitivity of the Olink assays requires a clean laboratory environment. Particles from the surroundings, such as dust, hair, saliva and skin flakes are common sources of contamination. The following recommendations are intended to simplify the workflow in the lab and improve data quality.

- Keep all consumables (tubes, pipette tips, PCR plates etc.) in closed bags or boxes, preferably in a closed storage unit.
- Always wear gloves, including when bringing reagents in and out from fridge or freezer.
- Change gloves when needed and always when going from pre-PCR to PCR operations.
- Always wear a long-sleeved lab coat.
- Use a separate lab coat for working with the PCR steps.
- Use separate rooms for pre-PCR and PCR operations. If this is not possible, keep separate benches.
- Use separate consumables and equipment for pre-PCR and PCR operations.

6. Preparation

6.1 Experimental design

It is important to calculate how many samples, replicates and controls that are needed to get the data you want from the study. When running more than one plate, Olink always recommends randomization of samples as good laboratory practice for all projects. Olink Target 48 utilizes a calibrator to determine the protein concentration and randomization of samples is recommended but not mandatory. Randomization refers to the random assignment of samples to all plates and wells included in the study.

It may be wise to consult a statistician or Olink Support (<u>support@olink.com</u>) prior to running the study. For more information, refer to the Olink white paper "Strategies for design of protein biomarker studies" available on the Olink website: <u>www.olink.com/downloads</u>.

6.2 Create Olink extension and pre-amplification program

1.	Create a	PEA program	on the thermal	cycler with t	the following	settings:
----	----------	-------------	----------------	---------------	---------------	-----------

Step	Temperature	Duration
Extension	50 °C	20 min
Hot start	95 °C	5 min
PCR cycle	95 °C	30 s
(17 cycles)	54 °C	1 min
	60 °C	1 min
Maintain the reaction at	10 °C	∞, hold

NOTE: Enable the heated lid function.

6.3 Bench setup

Before you start working it is important to know that you have everything you need and that all reagents are ready for pipetting. Do the following before you start the experiment:

- 1. Clean the lab bench, hood, racks and pipettes with 70% ethanol.
- 2. Bring out all reagents, consumables and samples needed for the specific lab step. Potential contamination from the surrounding environment is minimized by reducing physical movements in the lab.
- 3. Organize equipment, consumables and samples in the work station, in a way that enables clean work.
- 4. Label pipette boxes with column number to more easily monitor where you are on the plate.
- 5. Note that all consumables and reagents are single use only.

6.4 Time indications and limits

Step	Duration	Comment
Incubation setup	30-45 min	
Incubation	16-22 hours (overnight)	Keep incubation times consistent when running multiple IFCs for the same project (variation < 2 hours).
Extension	30 min	From preparing mix to start of PCR machine. IMPORTANT: A maximum of 5 minutes can pass between adding extension mix to the incubation plate until the start of the PCR. This is important for sensitivity.
Priming	10 min	 Prepare sample plate during priming of the IFC. NOTE: The primed IFC must be used within 60 minutes for the pressure to be maintained.
Load	1 hour 45 min	Start adding primers and samples to the primed IFC within 10 minutes of completed priming. Start the load script in the Olink Signature Q100 within 40 minutes from completion of priming. The time in the Signature Olink Q100 is 50 minutes.
qPCR	1 hour 30 min	After priming and pipetting the diluted PCR product and primers into the IFC, the Signature Q100 does the loading and qPCR steps automatically.

6.5 Sample preparation

Sample preparation recommendations are listed below:

- Use a 96-well PCR plate format. All plates need to be able to withstand -80 °C, be dry-ice resistant and easily resealable.
- Ensure that the samples are randomized, or in the order ready to be run and compatible with the plate layout.
- Ensure that each well is separately sealed using an adhesive film or individual seals.
- Clearly mark sample plates or tubes with a simple alphanumeric code that you can later identify (for example "A, B, C", "1, 2, 3", or "A1, A2, A3") using temperature-resistant labels or marker pen.
- Use unique sample identification names or numbers.

6.6 Plate layout

Below is an illustration of the plate layout. The sixth column consists of controls. The controls and calibrator are always included in the kit, and supplied by Olink.



7. Operation

7.1 Overview

The Olink protocol consists of three core steps. These are the Incubation step, the Extension and amplification step and the Detection step.



1. Incubation reaction

Duration: Overnight 16-22 hours Location: Pre-PCR room Needed for this step:

- Frw-probes
- Rev-probes
- Incubation Solution
- Sample Controls
- Negative Controls
- Calibrators

2. Extension and amplification Duration: 2 hours Location: Pre-PCR and PCR room Needed for this step:

- PEA Solution
- PEA Enzyme
- PEA Enhancer
- High Purity Water

3. Detection Duration: 2.5 hours

Location: PCR room Needed for this step:

- Detection Solution
- Detection Enzyme
- PCR Polymerase
- Primer plate
- Olink[®] 48.48 IFC for Protein Expression
- High Purity Water

7.2 Incubation reaction step

The Incubation reaction is where the antibody-pairs with attached DNA tags are added to the samples, and allowed to bind to their target proteins during an overnight incubation.

7.3 Prepare bench

Prepare the lab space with pipettes and consumables needed for this step and let the reagents and controls reach room temperature. For the Incubation reaction step, you will need:

- Incubation Solution
- Sample Controls
- Frw- and Rev-probes
- Negative Control
- Calibrator
- Two 8-well strips
- Prepared 96-well plate with samples in columns 1-5
- 1 (one) semi skirted 96-well plate

- 1 microcentrifuge tube (1-1.5 mL)
- 3 pipette tips (10 µL)
- 4 pipette tips (100/200 µL)
- 1 pipette tip (1000 µL)
- 56 F.A.S.T. Disposable Pipette Tips
- 2 adhesive films

Vortex and spin all reagents before use.

7.4 Incubation instruction

OTIP: Practice reverse pipetting before you start.

IMPORTANT: Always double check that you have the correct volume in ALL pipette tips.

- 1. Open the F.A.S.T. software and open protocol 'Olink_Target48Flex_Incubation_40 Samples'.
- 2. Thaw the samples. Vortex the sample plate using the MixMate at 2000 rpm for 30 seconds and spin down the liquid at 400-1000 x g, for 1 minute at room temperature.
- 3. Vortex the Sample Control, Negative Control and Calibrator and spin briefly. Add 4 µL of the Calibrator, Sample Control and Negative Control to an 8-well strip according to the following structure:



4. Prepare the Incubation Mix in a microcentrifuge tube according to the table below. Vortex and spin each reagent before adding it to the mix. Pipette the Incubation Solution carefully to avoid foaming. The Incubation Mix must be used within 45 minutes from preparation.

Incubation Mix	Per ½ 96-well plate (µL)
Olink® 1-48 plex Incubation Solution	176.0
Olink [®] Target 48 Frw-probes	22.0
Olink [®] Target 48 Rev-probes	22.0
Total	220.0

- 5. Vortex and spin down the Incubation Mix. Transfer 25 µL of the Incubation Mix to each well of a new 8-well strip.
- 6. Ensure there are no bubbles at the bottom of the control and Incubation Mix strips. Place the control strip in the strip tube adapter in column 12 and the Incubation Mix strip in column 1. This will be the Reagent Source.

- 7. Vortex the plate of samples using MixMate at 2000 rpm for 30 seconds and spin down the liquid at 400-1000 x g
- 8. Label a new 96-well semi skirted plate as 'Incubation plate' and place in adaptor.
- 9. Unseal and place the Reagent Source, Incubation Plate, Sample Source Plate, and tip box on the F.A.S.T. deck following the layout in the software. Tap **Start** to begin the protocol.

NOTE: The F.A.S.T. will first aliquot 3 μL of Incubation mix into the Incubation plate, followed by 1 μL of samples. Last it will pipette 1 μL of the controls from the Reagent Source into column 6.

10. Once the protocol finishes, seal the Incubation Plate thoroughly with an adhesive plastic film, spin at 400-1000 x g for 1 minute at room temperature.

IMPORTANT: It is critical to not vortex the Incubation Plates

- 11. Double check that all wells in the plate contain the same volume. Note any deviations. This information is needed to interpret the data in the statistical analysis step.
- 12. Incubate the Incubation Plate overnight at +4 °C for 16-22 hours in a refrigerator or cold room where the light is off when the door is closed.

NOTE: All wells must be properly sealed to avoid evaporation of the samples.

- 13. Prepare a tube of High Purity Water for the extension step the next day. Keep it at +4 °C overnight.
- 14. Thaw the PEA Solution over night at +4 °C, and place the PEA Enhancer at room temperature over night.

7.5 Extension and amplification step

In the morning of the following day, the extension and amplification steps take place. Unique DNA reporter sequences for each target protein are generated and pre-amplified using regular PCR.

Prepare bench

Prepare the lab space with pipettes and consumables needed for the Extension and amplification step. You will need:

- High Purity Water (+4 °C)
- PEA Solution (keep at room temperature for 15 minutes before use)
- PEA Enzyme

NOTE: Keep the PEA Enzyme in a freezing block or on ice.

- PEA Enhancer (at room temperature)
- 1 centrifuge tube (> 11 mL)
- 1 multichannel pipette reservoir
- 2 pipette tips (1000 µL)
- 2 pipette tips (100/200 µL)
- 8 multichannel pipette tips (200 µL)
- 1 adhesive film

7.5.1 Extension and amplification step instruction

IMPORTANT: Always double check that you have the correct volume in ALL pipette tips!

1. Allow the PEA Solution to reach room temperature. Vortex and spin down briefly before use. Make sure that no precipitation has occured. If there is precipitate, vortex and spin again.

- 2. Pre-heat the PCR machine to 50 $^\circ C$ and pause the PEA program.
- 3. Spin down the Incubation Plate at 400-1000 x g for 1 minute at room temperature.
- 4. Vortex and spin down the PEA Enzyme and PEA Enhancer briefly.
- 5. Prepare the Extension Mix in a 15 mL tube according to the table. Add the reagents in the specified order, starting with water. The Extension Mix must be used within 30 minutes from preparation.

Order	Extension Mix	Per ½ 96-well plate (µL)
1	High Purity Water (+4 °C)	4350
2	Olink [®] 1-48 plex PEA Enhancer	580
3	Olink [®] 1-48 plex PEA Solution	580
4	Olink [®] 1-48 plex PEA Enzyme	58
	Total	5568

- 6. Vortex the Extension Mix and pour it into a multichannel pipette reservoir.
- 7. Carefully remove the adhesive film from the Incubation Plate.

TIME SENSITIVE STEP: Perform steps 8-12 during 5 minutes.

8. Start a 5 minute timer and transfer 96 µL of Extension Mix to the upper parts of each of the well walls of column 1-6 in the Incubation Plate using **reverse pipetting**. Use the same pipette tips throughout the plate.

IMPORTANT: Do not let the tips come in contact with the content of a well.

9. Seal the plate with a new adhesive plastic film.

IMPORTANT: All wells must be properly sealed to avoid evaporation of the samples.

- 10. Use MixMate to vortex the plate thoroughly at 2000 rpm for 30 seconds to ensure that all wells are mixed before spinning it down.
- 11. Double check that all wells in the plate contain the same volume. Note any deviations. This information is needed to interpret the data in the statistical analysis step.
- 12. Take the Incubation Plate to the PCR room.
- 13. When the five minutes have passed, immediately place the Incubation Plate in the PCR instrument and resume the PCR PEA program. The program takes approximately 1 hour and 30 minutes.

NOTE: If the thermal cycler requires a silicon cover for plates covered with plastic film, use one to avoid evaporation.

 $\overline{\mathbf{Q}}$ TIP: You can start the preparations for the detection step during the last 10 minutes of the PEA program.

14. When the PCR PEA program is finished (~1 hour 55 minutes), continue to 7.6 Detection step.

SAFE STOPPING POINT: The extension products can be stored in the Incubation Plate for up to one week at +4 °C or up to four weeks at -20 °C in a refrigerator or freezer when the light is off when the door is closed.

7.6 Detection step

The final Detection step quantifies the DNA reporters for each biomarker using high throughput real-time qPCR on the Olink Signature Q100 system.

Prepare bench

For the priming of the IFC, you need:

- Olink[®] 48.48 IFC for Protein Expression
- Two syringes with control line fluid

For the Detection step:

- Detection Solution
- High purity water
- Detection Enzyme
- PCR Polymerase
- Primer Plate
- Extension products
- 1 (one) 8-well strip
- 1 (one) 96-well PCR plate
- 1 microcentrifuge tube (1-1.5 mL)
- 2 pipette tips (10 µL)
- 1 pipette tip (100/200 µL)
- 2 pipette tips (1000 µL)
- 152 F.A.S.T. Disposable Pipette Tips
- FORMULATRIX F.A.S.T. 96 well plate adapter
- FORMULATRIX F.A.S.T. strip tube adapter
- 2 adhesive films

Prime IFC



- 1. Unpack the syringes and actuate both check valves of the IFC with gentle pressure using a syringe. Ensure that the poppets of the check valves can move freely up and down.
- 2. Tilt the IFC and insert the syringe tip into the check valve opening. The tip should be fully inserted. Confirm that the valve is fully open with the O-ring seal pushed down and moved to the side.
- 3. Slowly inject one control line fluid syringe into each accumulator on the IFC. Make sure that the fluid flows away from the accumulator, otherwise a vacuum is formed and the oil bubbles up.

- 4. Remove the film from the bottom of the IFC.
- 5. On the Home screen of the Olink Signature Q100 instrument, tap Target 48. Place the IFC containing control line fluid and the interface plate on the drawer. Align the notched corner of the IFC to the notch on the drawer and face the barcoded edges of the IFC and interface plate forward. Tap Close Drawer. On the Ready to start chip preparation screen, tap Start. The screen shows the remaining time. For further details, refer to the Olink[®] Signature Q100 User Manual.

NOTE: Prepare the Detection Mix during the priming of the IFC.

Detection step instruction

 $\stackrel{\frown}{\longrightarrow}$ IMPORTANT: Always double check that you have the correct volume in ALL pipette tips.

- 1. Open the F.A.S.T. software and open the program named 'Olink_Target48Flex_Detection_40 Samples_Left Side'.
- 2. Remove the Incubation plate (Extension products) from the PCR instrument. Vortex and spin down the liquid.
- 3. Double check that all wells in the plate contain the same volume. Note any deviations. This information is needed to interpret the data in the statistical analysis step.
- 4. Thaw the Primer Plate and Detection Solution. Vortex and spin down the Detection Solution and Primer Plate. Keep the Detection Enzyme and PCR Polymerase in a freezing block or on ice.
- 5. Prepare the Detection Mix in a microcentrifuge tube. Vortex and spin down the enzymes before adding them to the mix.

Detection Mix	Per ½ 96-well plate (µL)
Olink® 1-48 plex Detection Solution	275.0
High purity water	116.0
Olink® 1-48 plex Detection Enzyme	3.9
Olink [®] 1-48 plex PCR Polymerase	1.5
Total	396.4

- 6. Vortex the Detection Mix and spin briefly. Transfer 47 µL of the mix to each well of an 8-well strip, ensure there are no bubbles at the bottom of the wells, and place in the strip tube adapter for the Reagent Source.
- 7. Label a new 96-well plate 'Detection Plate'.
- 8. Carefully remove the adhesive film from the Incubation Plate.
- 9. Remove the primed IFC from the Olink Signature Q100.

IMPORTANT: Do not touch the interface plate actuating pins once affixed to the IFC, since that could result in depressurizing of the accumulators.



Unseal and place the Reagent Source, Incubation Plate, Detection Plate, IFC, Primer Plate, and tip box on the F.A.S.T. deck following the layout in the software.

NOTE: The IFC should be oriented so that the cut corner of the chip is placed on the upper right side.

NOTE: The F.A.S.T. will aliquot first detection mix into the 'Detection Plate', followed by extension product from the Incubation Plate, then pause.



Primer Plate

?: The plate with extension products can be saved up to one week at +4 °C or up to four weeks at -20 °C in a refrigerator or freezer where the light is off when the door is closed.

NOTE: The F.A.S.T. will load 5 μ L primers into the IFC plate, eight wells at a time and then 5 μ L of the samples from the Detection Plate into the IFC Plate, eight wells at a time.

- 10. Vortex the Detection Plate using the MixMate at 2000 rpm for 30 seconds and spin down the liquid at 400-1000 x g, for 1 minute at room temperature.
- 11. Double check that all wells in the Detection Plate contain the same volume. Note any deviations. This information is needed to interpret the data in the statistical analysis step.
- 12. Unseal the Detection Plate and place on the tray according to the layout. Tap **Confirm** to continue with the protocol.

NOTE: The total loading of the IFC takes approx. 10 minutes. In the previous steps, the primers were loaded to the left and samples to the right on the Olink[®] 48.48 IFC for Protein Expression. Here follows an overview illustration of the loading:



Primers

Samples



- 13. When loading is finished, inspect the wells and remove any bubbles using a syringe needle. Change needle between wells to avoid contamination.
- 14. Use a piece of adhesive tape to remove potential dust from the top of the middle section of the IFC. Let the sticky part lightly touch the surface of the IFC.
- 15. On the Pipette samples and primers onto the chip screen on Olink Signature Q100, tap **Open drawer**.

P IMPORTANT: Do not touch the interface plate actuating pins once affixed to the IFC, since that could result in depressurizing of the accumulators.



Place the correct interface plate on the IFC, put the IFC with the interface plate in the open drawer of the Olink Signature Q100. Make sure that the Interface plate is placed properly on the IFC and in level with the drawer. Align the notched corner of the IFC with the notch on the drawer and face the barcoded edges of the IFC and interface plate forward. Tap **Close Drawer**. On the Add run details screen, confirm the Chip ID, then enter either Run Name or Run Notes. Tap **Start Run**. The screen shows the remaining time.

16. When the run is complete, remove the IFC and interface plate from the instrument.

8. StandardBioTools® BioMark[™] system

The Olink Target 48 panels can be run both on the Olink Signature Q100 instrument as well as on the Standard BioTools BioMark system. If using the BioMark system the additional requirments and operational steps of this chapter should be followed.

8.1 Additional requirements: Analysis software

The BioMark Data Collection software needs to be installed on the BioMark computer, and Real-Time PCR Analysis software is required for the analysis of Olink data in the NPX[™] Signature software.

8.2 Preparation - Prepare files

8.2.1 Enable script on Juno System for loading of Olink® 48.48 IFC for Protein Expression

To load the Olink[®] 48.48 IFC for Protein Expression on the Juno System, the correct script must be enabled.

NOTE: The script for loading on the Juno System is enabled by default, 48.48 CS Load Mix.

- 1. Log in as administrator, tap Admin below the tabs (boxes).
- 2. On the Tools screen, tap Preferences.
- 3. Select Enabled under Advanced Scripts.
- 4. On the IFC control tab, select Load Mix 48.48 GT.

8.2.2 Replace definition file

To run the Olink[®] 48.48 IFC for Protein Expression on the BioMark HD with the Olink program, the definition file needs to be replaced.

NOTE: This should not be performed while the BioMark is running a program.

- 1. Use the BioMark computer to find the IFC definition file M48CS2ChipDefinition.xmx from the <u>Olink software</u> <u>download page</u>.
- 2. Right-click and click **Save as**, and navigate to C:/*Program files/Fluidigm/BioMarkDataCollection/ApplicationData/ Chip Definitions*.
- 3. Replace the current M48CS2ChipDefinition.xmx file with the M48CS2ChipDefinition.xmx file from the <u>Olink</u> <u>software download page</u>.

NOTE: This IFC definition file retains all the original functionality of the replaced file. In addition it has added functionality for the Olink Target 48 assay.

4. Restart the BioMark Data Collection Software.

8.2.3 Import the 48x48 Olink Protein Expression 40 cycles program

- 1. On the BioMark computer, go to My Computer and select C:.
- 2. Navigate to Program files/Fluidigm/BioMarkDataCollection/ApplicationData/Protocols.

- 3. Create a new folder, and name it *Olink*.
- 4. Find the 48x48 Olink Protein Expression 40 cycles.pcl file on the <u>Olink software download page</u>. Right-click and click **Save as**.
- 5. Navigate to C:/*Program files/Fluidigm/BioMarkDataCollection/ApplicationData/Protocols/Olink* and save the 48x48 Olink Protein Expression 40 cycles.pcl file in the *Olink* folder.

8.2.4 Create the Olink project setting on the BioMark

- 1. Open the BioMark Data Collection Software on the BioMark computer.
- 2. Go to Settings>Manage Project Settings.
- 3. Click Add New Project Setting and name the project 48x48 Olink Protein Expression 40 cycles .
- 4. Make the following settings:

Variable	Setting
Application	Gene Expression
Passive Reference	ROX (5-Carboxy-x-Rhodamine)
Assay	Single probe
Probes	FAM-MGB

- 5. Under **Naming and Location, Chip Run Directory**, click **Browse**. Navigate to the folder where you want to save the data files.
- 6. To upload the protocol, click **Browse for Protocol** under **Thermal Protocol** section and select the 48x48 Olink Protein Expression 40 cycles protocol.
- 7. Save and Close and then click Close again.
- 8. Select the **Project Setting** 48x48 Olink Protein Expression 40 cycles in the BioMark Data Collection starting window

 $\dot{\mathbf{Q}}$ TIP: Select the Remember Project Setting checkbox to make it the default project setting.

The 48x48 Olink Protein Expression 40 cycles program includes the following steps:

Step	Temperature	Duration
Pre-heat	50°C	120 s
Hot start	95°C	300 s
PCR cycle (40 cycles)	95°C	15 s
	60°C	60 s

8.2.5 Olink extension and pre-amplification program settings

Step	Temperature	Duration
Extension	50 °C	20 min
Hot start	95 °C	5 min
PCR cycle	95 °C	30 s
(17 cycles)	54 °C	1 min
	60 °C	1 min
Maintain the reaction at	10 °C	∞, hold

Create a PEA program on the thermal cycler with the following settings:

9. When preparation is complete, tap rightarrow (Eject) in the bottom right corner to open the drawer.

Continue to 6.3 Bench setup.

8.3 Operation

8.3.1 Detection step

Prime IFC

After removing the film from the bottom of the IFC (step 4), place the IFC, with barcoding facing you, in the IFC Controller. On the touch-screen, first press **Load Chip** or **LOAD** and then select the program *48x48CS Chip Prime* for IFC controller MX and *Prime 48.48 GT* for Juno System. Press **Prime** or **RUN**. The program takes approximately 10 minutes.

Continue to Detection step instruction.

Detection step instruction

When finished with step 22, continue with:

- Place the IFC with its barcoding facing you in the IFC Controller. Select Load Chip or LOAD and the script to run. Use the Load Mix (124X) on the MX controller or the Load Mix 48.48 GT on the Juno System, followed by Run Script or RUN to load the assay and sample mixes into the central part of the IFC. The loading program takes approximately 50 minutes. See the manufacturer's instructions for more details.
- 2. When the IFC loading is completed, eject the IFC from the IFC controller.
- 3. Use a piece of adhesive tape to remove dust from the top of the middle section of the IFC. Let the sticky part lightly touch the surface of the IFC.
- 4. Load the IFC in the BioMark Reader with the barcode facing outwards and start the 48x48 Olink Protein Expression 40 cycles protocol.

The BioMark program takes approximately 1 hour and 30 minutes.

9. Other resources

9.1 Olink FAQs

The answers to the most common questions asked by our customers can be found on the Olink website: <u>www.olink.com/</u><u>faq</u>.

9.2 Data analysis troubleshooting

Regarding issues detected during the quality control of the data or statistical analysis, see the troubleshooting chapter in the NPX[™] Signature User Manual for solutions. The manual is available from the Olink website: <u>www.olink.com/</u><u>downloads</u>.

9.3 Lab instruction video

Learn from our experienced Analysis Service team how to run Olink panels by watching the <u>Olink Lab Instruction video</u> on the Olink youtube channel.

9.4 PEA technology video

For an animated description of how our innovative dual recognition, DNA-coupled methodology provides exceptional readout specificity, watch the <u>PEA overview</u> on the Olink youtube channel.

10. Revision history

Version	Date	Description
1.1	2024-03-14	Added notes about not touching the interface plate actuating pins once affixed to the IFC, in section 7.6.
1.0	2023-10-26	New document including instructions for the use of Formulatrix F.A.S.T.

www.olink.com

© 2024 Olink Proteomics AB.

Olink products and services are For Research Use Only and not for Use in Diagnostic Procedures.

All information in this document is subject to change without notice. This document is not intended to convey any warranties, representations and/or recommendations of any kind, unless such warranties, representations and/or recommendations are explicitly stated.

Olink assumes no liability arising from a prospective reader's actions based on this document.

OLINK, NPX, PEA, PROXIMITY EXTENSION, INSIGHT and the Olink logotype are trademarks registered, or pending registration, by Olink Proteomics AB. All third-party trademarks are the property of their respective owners.

Olink products and assay methods are covered by several patents and patent applications https://www.olink.com/patents/.